

Testing the virucidal activity of "Liquid Guard"

Examination of test surfaces equipped with a virucidal active coating using a praxis-near carrier test system following the RKI-Richtlinie (1995) as well as ISO 21702:2019 against the *Transmissible Gastroenteritis Virus (TGEV-Coronavirus)* - Test run S1 dated 11./12.03.2020

Short report: screening test S2

by

PD Dr. Olaf Thraenhart and Dr. Christian Jursch

Test period: in March 2020

Principal: Nano-Care Deutschland AG
Alfred Nobel-Straße 10
D-66793 Saarwellingen, Germany

Principal: Nano-Care UK (Signo-Nanocare UK Ltd)
PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Products:

- Test surfaces: *Leneta*[®] foil, with the dimensions of 1,6 cm x 6 cm
- 1. test item: test surfaces coated on one side with Liquid Guard (containing the active component[s])
- 2. test item: uncoated test surfaces (or coated w/o the active component[s])

Test parameter:

- Test conditions: T = 25 °C and 90 % r.LF
- Protein load: no additional protein load; the virus material (cell culture supernatant) was spread onto the surface(s) w/o any further manipulation/alteration
- Volume to square ratio: 25 µL/cm²
- Virus suspension covered with foil (LDPE, 50 µm) with the dimensions 1,2 x 5 cm (6 cm²)
- Incubation: 1h, 8h and 24h in a climate chamber KBF 115 (Fa Binder)

Test system:

- Transmissible Gastroenteritis Virus of Swine (TGEV-Coronavirus); strain: Toyama 36 [used in test as the model virus for SARS-CoV]
(Origin: Virusbank of the Friedrich Löffler-Institute, Insel Riems, Germany)
- ST75/2 cells (foetal testis cells of swine)
(Origin: Robert Koch-Institute, Berlin, Germany)

Test procedure:

- The test was performed following a. RKI-Richtlinie (1995) as well as b. ISO 21702:2019
- Test principle: quantitative virucidal carrier test at T = 25 °C and 90 % r.LF (climate chamber)
- the test was performed w/o (additional) protein load

Tab. 1: Product samples tested

No.	Product (s)	Storage conditions ¹
#1	Test item / coated with <u>Liquid Guard</u> (containing the virucidal active component(s) / „test sample“)	at RT
#2	Test item / uncoated (or coated w/o the virucidal active component(s) / „control sample“)	at RT

¹ = access limited

Test results:

Observations:

- The test surfaces were largely wettable by the aqueous virus suspension; thus, a more or less uniform liquid film could be produced by using glass spatulas.
- After covering the virus with the LDPE foil, the virus material remained stable as a film over the entire observation period and did not dry out. However, a volume reduction was recorded.

Tab. 2.1: Virus control (Virus titration by limiting dilution)

Sample	VK-1a	VK-1b	VK-2a	VK-2b	VK-3a	VK-3b
	Virus control / 1 h		Virus control / 8 h		Virus control / 24 h	
Titer/Test vol. (lg ID ₅₀)	4,2	4,8	4,05	3,9	2,25	2,85
av. virus titer ± K (95%)¹	5,50 ± 0,37 / 1 mL		4,98 ± 0,35 / 1 mL		3,55 ± 0,37 / 1 mL	

¹ = Calculation of the virus titer and its 95% confidence interval according to EN14476

Tab. 2.2: Virus inactivation (Virus titration by limiting dilution)

Sample	In-1a	In-1b	In-2a	In-2b	In-3a	In-3b
	Inactivation / 1 h		Inactivation / 8 h		Inactivation / 24 h	
Titer/Test vol. (lg ID ₅₀)	3,6	3,45	1,35	1,2	≤ 0,30	≤ 0,30
av. virus titer ± K (95%)¹	4,53 ± 0,22 / mL		2,28 ± 0,29 / mL		≤ 1,30 / mL	
Reduction² (lg ID₅₀ ± K [95%])	0,97 ± 0,43		2,70 ± 0,46		≥ 2,25 ± 0,37	

¹ = Calculation of the virus titer and its 95% confidence interval according to EN14476

² = Virus reduction: lg ID₅₀ of virus input (virus control) minus lg ID₅₀ of sample (at the given time point)

Virus inactivation: (cf. Tab. 2)

- When the virus material is distributed onto a surface a certain virus titer reduction could be observed with almost all viruses. This is driven by time and do also occur without any other influence. This is also true for the test virus used in the present testing. After presentation over 8 h and 24 h on the test surface a titer reduction of 0,5 Log was evident after 8 h and about 2 Log after 24 h (cf. tab. 2.1). It should be noted, however, that this reduction can be judged as very low when compared to 1). the general tenacity of coronaviruses and b). other viruses (even non-enveloped viruses).
- In order to assess the virus inactivating capacity of the coating under test as a single factor an individual virus input control was analysed at each time point tested. With the amount of input virus at a given time point (cf. tab. 2.1) and with the correspondent amount of remaining test virus (cf. tab. 2.2) the virus reduction factor can be determined.
- After the incubation time was due and under the test conditions specified above the virus reduction factor associated with the coating containing the active component amounted to RF = 0,97 ± 0,43 after 1 h, to RF = 2,70 ± 0,46 after 8 h and to RF ≥ 2,25 ± 0,37 after 24 h (cf. Tab. 2.2). It should be noted that after 24 h no residual test virus was detectable.

Conclusions:

- The virus film applied on the test items and covered with the LDPE-foil was stable over the entire observation period. This means that the virus film remained in the liquid state even at the end of the longest exposure time (24 h) and was not dried. Thus, a continuous contact between the virus material and the surface of the test carrier was ensured all over the observation period and a distribution of the virus material in the liquid phase driven by diffusion was given.
- After $t = 1$ h a virus reduction of 0,97 Log was recorded (corresponding to 90 % of inactivation) and after $t = 8$ h the virus reduction amounted to $RF = 2,7$ (corresponding to 99,8 % of inactivation). Due to technical reasons demonstration of the virus reduction was limited to $RF \geq 2,25$ after 24 h.
- The data obtained allow the conclusion that there is a virus reduction that can be attributed to the coating containing the active component(s). With the present testing a good virus inactivating activity of the virucidal coating under test was demonstrated against the TGEV-Coronavirus (as the model virus for the SARS-CoV).
- It should also be mentioned that the conditions of ISO 21702 provide for a higher incubation temperature than that used in S1 (25 vs. 21 ° C).
- The virus reduction obtained with $t = 8$ h suggests that at the incubation time $t = 24$ h a higher virus reduction is evident than could be demonstrated with the endpoint titration method. Here, virus titer determination using the *Large Volume Plating (LVP)* can possibly provide an improved statement.

Luckenwalde, 20th of March 2020



Dr. Ch. Jursch
(GF und Laborleiter Eurovir)

Testing the virucidal activity of "Liquid Guard"

Examination of test surfaces equipped with a virucidal active coating using a praxis-near carrier test system following the RKI-Richtlinie (1995) as well as JIS Z 2801 (2010) against *Influenza A Virus (H1N1)* - Test run S1 dated 20./21.01.2020

Short report: screening test S1

by

PD Dr. Olaf Thraenhart and Dr. Christian Jursch

Test period: in January 2020

Principal: Nano-Care Deutschland AG
Alfred Nobel-Straße 10
D-66793 Saarwellingen, Germany

Principal: Nano-Care UK (Signo-Nanocare UK Ltd)
PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Products:

- Test surfaces: *Leneta*[®] foil, with the dimensions of 1,6 cm x 6 cm
- 1. test item: test surfaces coated on one side with Liquid Guard (containing the active component[s])
- 2. test item: uncoated test surfaces (or coated w/o the active component[s])

Test parameter:

- Test conditions: T = 21 °C and 60 % r.LF
- Protein load: no additional protein load; the virus material (cell culture supernatant) was spread onto the surface(s) w/o any further manipulation/alteration
- Volume to square ratio: 20 µL/cm²
- Virus suspension covered with foil (LDPE, 50 µm) with the dimensions 1,2 x 5 cm (6 cm²)
- Incubation: 1h, 8h and 24h in a climate chamber KBF 115 (Fa Binder)

Test system:

- Influenza A Virus; H1N1; strain: New Caledonia
(Origin: Chiron Behring, Marburg, Germany)
- MDCK-cells (kidney cells from African green monkey [*Cercopithecus aethiops*])
(Origin: Robert Koch-Institut, Berlin, Germany)

Test procedure:

- The test was performed following a. RKI-Richtlinie (1995) as well as b. JIS Z 2801 (2010)
- Test principle: quantitative virucidal carrier test at T = 21 °C and 60 % r.LF (climate chamber)
- the test was performed w/o (additional) protein load

Tab. 1: Product samples tested (as received at 13.01.2020)

No.	Product (s)	Storage conditions ¹
#1	Test item / coated with <u>Liquid Guard</u> (containing the virucidal active component(s) / „test sample“)	at RT
#2	Test item / uncoated (or coated w/o the virucidal active component(s) / „control sample“)	at RT

¹ = access limited to the personnel of Eurovir

Test results:

Observations:

- The test surfaces were largely wettable by the aqueous virus suspension; thus, a more or less uniform liquid film could be produced by using glass spatulas.
- After covering the virus with the LDPE foil, the virus material remained stable as a film over the entire observation period and did not dry out. However, a volume reduction was recorded.

Tab. 2.1: Virus control (Virus titration by limiting dilution)

Sample	VK-1a	VK-1b	VK-2a	VK-2b	VK-3a	VK-3b
	Virus control / 1 h		Virus control / 8 h		Virus control / 24 h	
Titer/Test vol. (lg ID ₅₀)	3,15	3,3	3,45	3,15	2,7	3,15
av. virus titer ± K (95%)¹	3,23 ± 0,36 / 100 µL		3,30 ± 0,33 / 100 µL		2,93 ± 0,34 / 100 µL	

¹ = Calculation of the virus titer and its 95% confidence interval according to EN14476

Tab. 2.2: Virus inactivation (Virus titration by limiting dilution)

Sample	In-1a	In-1b	In-2a	In-2b	In-3a	In-3b
	Inactivation / 1 h		Inactivation / 8 h		Inactivation / 24 h	
Titer/Test vol. (lg ID ₅₀)	3,45	3,15	2,4	2,4	1,2	1,2
av. virus titer ± K (95%) ¹	3,30 ± 0,32		2,40 ± 0,29		1,20 ± 0,33	
Reduction² (lg ID₅₀ ± K [95%])	-0,07 ± 0,48		0,90 ± 0,44		1,73 ± 0,47	

¹ = Calculation of the virus titer and its 95% confidence interval according to EN14476

² = Virus reduction: lg ID₅₀ of virus input (virus control) minus lg ID₅₀ of sample (at the given time point)

Virus inactivation: (cf. Tab. 2)

- When the virus material is distributed onto a surface a certain virus titer reduction could be observed with almost all viruses. This is driven by time and do also occur without any other influence. This is also true for the test virus used in the present testing. After presentation over 24 h on the test surface a titer reduction of 0,3 Log was evident (cf. tab. 2.1). It should be noted, however, that this reduction can be judged as very low when compared to 1). the general tenacity of influenza virus and b). other viruses (even non-enveloped viruses).
- In order to assess the virus inactivating capacity of the coating under test as a single factor an individual virus input control was analysed at each time point tested. With the amount of input virus at a given time point (cf. tab. 2.1) and with the correspondent amount of remaining test virus (cf. tab. 2.2) the virus reduction factor can be determined.
- After the incubation time was due and under the test conditions specified above the virus reduction factor associated with the coating containing the active component amounted to RF = -0,07 ± 0,48 after 1 h, to RF = 0,90 ± 0,44 after 8 h and to RF = 1,73 ± 0,47 after 24 h (cf. Tab. 2.2).

Conclusions:

- The virus film applied on the test items and covered with the LDPE-foil was stable over the entire observation period. This means that the virus film remained in the liquid state even at the end of the longest exposure time (24 h) and was not dried. Thus, a continuous contact between the virus material and the surface of the test carrier was ensured all over the observation period and a distribution of the virus material in the liquid phase driven by diffusion was given.
- The data obtained allow the conclusion that there is a virus reduction that can be attributed to the coating containing the active component(s).
- The virus reduction rate progresses rather slowly over the observation period. No virus inactivation was detectable after a contact time of 1 hour and after 8 hours the virus reduction was approximately 1 Log (corresponding to a virus reduction of approximately 90%). After 24 hours virus reduction reached approximately 2 Log (corresponding to a reduction of approximately 99%).
- The observed virus-inactivating effect of the coating (containing the active component[s]) was determined using the *influenza A virus* as the test virus. This virus is in general considered to be inactivated easily, even when compared with other enveloped virus. This means that the observed virus inactivation capacity of the tested coating, as obtained with *influenza A virus*, cannot be transferred necessarily to other viruses. This also applies to other enveloped viruses.

Luckenwalde, 4th of March 2020



Dr. Ch. Jursch

(GF and Laboratory manager of Eurovir)

Work Order	3109.4W
Setup-Code	180612-10290-2801-01



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and Effectiveness

Test Object:

Coated Leneta-Foil vs. Aspergillus niger DSM 1988

Work Order	3109.4W
Setup-Code	180612-10290-2801-01

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3109.4W

Test object: Coated Leneta-Foil vs. Aspergillus niger DSM 1988

Sample description: Coated Leneta Foil

Date of receipt of sample: 14.09.2018

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: Aspergillus niger DSM 1988

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 180612-10290-2801-01

Sample material: Leneta-Foil

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 14.9.2018
Recipient: Nano-Care Deutschland AG

Laboratory Director:

Harald Gerauer, Laboratory Director
QualityLabs BT GmbH

Released:

Markus Zehe, Managing Director
QualityLabs BT GmbH

Work Order	3109.4W
Setup-Code	180612-10290-2801-01

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3109.4W
Setup-Code	180612-10290-2801-01

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3109.4W
Setup-Code	180612-10290-2801-01

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning	Isopropanol	-

References to deviations, preincubations, special test conditions

NONE

Work Order	3109.4
Setup-Code	180612-10290-2801-01

Test Results

Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1 Leneta-Foil P121-10 (Reference)	102900806180001	2,5 x 10 ⁴	2,7 x 10 ⁴	2,6 x 10 ⁴	2,1 x 10 ⁴	2,5 x 10 ⁴	2,7 x 10 ⁴	-	-
2 Liquid Guard clean + primer + wipe	102900806180002				6,4 x 10 ³	6,1 x 10 ³	3,5 x 10 ³	78,12	0,66
3 Liquid Guard pro + wipe	102900806180003				1,9 x 10 ⁴	6,1 x 10 ³	2,5 x 10 ⁴	0,02	0,02

*see "Interpretation of Results", page 6

Test strain	<i>Aspergillus niger DSM1988</i>
Initial cell count inoculum / cm ²	1.25 x 10⁴
Initials of the editor	OS
Measurement ended on	17.07.2018

Work Order	3109.4
Setup-Code	180612-10290-2801-01

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Shendi _____

Crosschecked: Mr. Zehe _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3250
Setup-Code	181218-10290-2801-01



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

Coated Leneta-Foil versus Clostridioides difficile DSM 27543

Work Order	3250
Setup-Code	181218-10290-2801-01

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3250

Test object: Coated Leneta-Fiol versus *Clostridioides difficile* DSM 27543

Sample description: coated foil

Date of receipt of sample: 2018-Oct-22

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: *Clostridioides difficile* DSM 27543

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 181218-10290-2801-01

Sample material: n.b.

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 2018-Dez-18
Recipient: Nano-Care Deutschland GmbH

Laboratory Director:

Harald Gerauer, Laboratory Director
QualityLabs BT GmbH

Released:

Markus Zehe, Managing Director
QualityLabs BT GmbH

Work Order	3250
Setup-Code	181218-10290-2801-01

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3250
Setup-Code	181218-10290-2801-01

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3250
Setup-Code	181218-10290-2801-01

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning	-	-

References to deviations, preincubations, special test conditions

NONE

Work Order	3250
Setup-Code	181218-10290-2801-01

Test Results

Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1 Reference sample	102902310180011	1.1 x 10 ³	8.3 x 10 ²	6.0 x 10 ²	2.3 x 10 ²	1.5 x 10 ²	1.5 x 10 ²		Reference
2 Liquid Guard 2,9%	102902310180012				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	3.45

*see "Interpretation of Results", page 6

Test strain	<i>Clostridioides difficile</i> DSM 27543
Initial cell count inoculum / cm ²	1.25 x 10 ⁴
Initials of the editor	MZ
Measurement ended on	Dec-12-2018

Work Order	3250
Setup-Code	181218-10290-2801-01

Comments on test objects

NONE

Interpretation of the results based on the measurements

Due to the test germ used *Clostridioides difficile* DSM 27543 exclusively grows anaerobic, under exclusion of air, the test specimens, and the agar plates were incubated after inoculation and incubation at 37 ° C in anaerobic pots. However, relatively low growth was seen on the blank immediately after inoculation (t0) and after incubation (t24). But there was a significant reduction in germs due to the Liquid Guard 2.9% coated Leneta film (102902310180012) compared to the blank sample (102902310180011).

Editor: Mr. Zehe _____

Crosschecked: Mr. Shendi _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3223.3_Rev3
Setup-Code	190109-10290-2801-01



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

*Coated leneta foil against Enterococcus hirae DSM3320
ATCC10541 after one year of artificial aging*

Work Order	3223.3_Rev3
Setup-Code	190109-10290-2801-01

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3223.3_Rev3

Test object: Coated leneta foil against *Enterococcus hirae* DSM3320 ATCC10541 after one year of artificial aging

Sample description: painted plastic films

Date of receipt of sample: 2018-Oct-22

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: *Enterococcus hirae* DSM3320 ATCC10541

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 190109-10290-2801-01

Sample material: Leneta foil

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 2019-Feb-27
Recipient: Nano-Care Deutschland GmbH
 replaces the test report from the 20th Jan. 2019

Laboratory Director: _____
 Harald Gerauer, Laboratory Director
 QualityLabs BT GmbH

Released: _____
 Markus Zehe, Managing Director
 QualityLabs BT GmbH

Work Order	3223.3_Rev3
Setup-Code	190109-10290-2801-01

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3223.3_Rev3
Setup-Code	190109-10290-2801-01

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3223.3_Rev3
Setup-Code	190109-10290-2801-01

References to Testconditions

Test conditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning		-

References to deviations, preincubations, special test conditions

The samples have been stored for 56 days at 50 °C, which corresponds to an artificial aging of 1 year according to ASTM F 1980.

Work Order	3223.3_Rev3
Setup-Code	190109-10290-2801-01

Test Results

Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1 Reference sample	102902310180011	1.1 x 10 ⁵	9.4 x 10 ⁴	1.2 x 10 ⁵	1.3 x 10 ⁵	1.4 x 10 ⁵	7.6 x 10 ⁴		Reference
2 Liquid Guard 2,9%	102902310180012				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Enterococcus hirae</i> DSM3320 ATCC10541
Initial cell count inoculum / cm²	1.25 x 10⁴
Initials of the editor	GM
Measurement ended on	17-01-2019

Work Order	3223.3_Rev3
Setup-Code	190109-10290-2801-01

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Shendi_____

Crosschecked: Mr. Mannala_____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3109.1
Setup-Code	180611-10290-2801-01



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

Coated Leneta-Foil vs. E. coli DSM 1576

Work Order	3109.1
Setup-Code	180611-10290-2801-01

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3109.1

Test object: Coated Leneta-Folie vs. E. coli DSM 1576

Sample description: lackierte Kunststofffolien

Date of receipt of sample: 08.06.2018

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: Escherichia coli DSM1576 ATCC8739 ISML CC 02/023

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 180611-10290-2801-01

Sample material: Leneta-Folie

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 14.9.2018
Recipient: Nano-Care Deutschland AG

Laboratory Director:

Harald Gerauer, Laboratory Director
QualityLabs BT GmbH

Released:

Markus Zehe, Managing Director
QualityLabs BT GmbH

Work Order	3109.1
Setup-Code	180611-10290-2801-01

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3109.1
Setup-Code	180611-10290-2801-01

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3109.1
Setup-Code	180611-10290-2801-01

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning	Isopropanol	-

References to deviations, preincubations, special test conditions

NONE

Work Order	3109.1
Setup-Code	180611-10290-2801-01

Test Results

	Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1	Leneta-Folie P121-10 (Reference)	102900806180001	9.0 x 10 ⁴	1.1 x 10 ⁵	1.1 x 10 ⁵	5.1 x 10 ⁵	4.1 x 10 ⁵	5.3 x 10 ⁵	-	-
2	Liquid Guard clean + primer + wipe	102900806180002				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4
3	Liquid Guard pro + wipe	102900806180003				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Escherichia coli</i> DSM1576 ATCC8739 ISML CC 02/023
Initial cell count inoculum / cm ²	1.25 x 10 ⁴
Initials of the editor	MZ
Measurement ended on	Jun-15-2018

Work Order	3109.1
Setup-Code	180611-10290-2801-01

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Zehe _____

Crosschecked: Mr. Shendi _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3382
Setup-Code	190710-10290-2801-01



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

Liquid Guard 2,9 % versus Escherichia coli DSM22312

Work Order	3382
Setup-Code	190710-10290-2801-01

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3382

Test object: Liquid Guard 2,9 % versus *Escherichia coli* DSM22312

Sample description: Coated foil

Date of receipt of sample: 2019-Jul-03

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: *Escherichia coli* DSM22312

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 190710-10290-2801-01

Sample material: n.b.

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 2019-Jul-18
Recipient: Nano-Care Deutschland GmbH

Laboratory Director:

Harald Gerauer, Laboratory Director
QualityLabs BT GmbH

Released:

Markus Zehe, Managing Director
QualityLabs BT GmbH

Work Order	3382
Setup-Code	190710-10290-2801-01

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3382
Setup-Code	190710-10290-2801-01

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3382
Setup-Code	190710-10290-2801-01

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning		-

References to deviations, preincubations, special test conditions

NONE

Test Report JIS Z 2801:2012 (Mod)

Work Order	3382
Setup-Code	190710-10290-2801-01

Test Results

Sample Name	Sample Code	t_0 (cells/cm ²)			t_{24} (cells/cm ²)			Reduction [%]	Log Reduction
1 Reference sample	102901007190019	7.7 x 10 ⁴	10.0 x 10 ⁴	8.9 x 10 ⁴	8.6 x 10 ⁴	9.9 x 10 ⁴	1.1 x 10 ⁵		Reference
2 Liquid Guard 2,9%	102901007190020				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Escherichia coli</i> DSM22312
Initial cell count inoculum / cm²	1.25 x 10 ⁴
Initials of the editor	MZ
Measurement ended on	Jul-12-2019

Work Order	3382
Setup-Code	190710-10290-2801-01

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Zehe _____

Crosschecked: Mr. Mannala _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3109.3
Setup-Code	180611-10290-2801-03



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

Coated Leneta Foil vs. Klebsiella pneumoniae DSM 6135

Work Order	3109.3
Setup-Code	180611-10290-2801-03

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3109.3

Test object: Coated Leneta Foil vs. Klebsiella pneumoniae DSM 6135

Sample description: Coated Foil

Date of receipt of sample: 08.06.2018

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: Klebsiella pneumoniae DSM6135

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 180611-10290-2801-03

Sample material: Leneta Foil

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 14.9.2018
Recipient: Nano-Care Deutschland AG

Laboratory Director:

 Harald Gerauer, Laboratory Director
 QualityLabs BT GmbH

Released:

 Markus Zehe, Managing Director
 QualityLabs BT GmbH

Work Order	3109.3
Setup-Code	180611-10290-2801-03

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3109.3
Setup-Code	180611-10290-2801-03

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3109.3
Setup-Code	180611-10290-2801-03

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning	Isopropanol	-

References to deviations, preincubations, special test conditions

NONE

Work Order	3109.3
Setup-Code	180611-10290-2801-03

Test Results

	Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1	Leneta-Folie P121-10 (Reference)	102900806180001	2.1 x 10 ⁴	3.6 x 10 ⁴	8.0 x 10 ⁴	2.7 x 10 ⁵	3.8 x 10 ⁵	2.7 x 10 ⁵	-	-
2	Liquid Guard clean + primer + wipe	102900806180002				6.4 x 10 ²	7.5 x 10 ¹	< 1.0 x 10 ¹	99.92	3.11
3	Liquid Guard pro + wipe	102900806180003				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Klebsiella pneumoniae DSM6135</i>
Initial cell count inoculum / cm ²	1.25 x 10 ⁴
Initials of the editor	MZ
Measurement ended on	Mar-12-20xx

Work Order	3109.3
Setup-Code	180611-10290-2801-03

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Zehe _____

Crosschecked: Mr. Shendi _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3109.6
Setup-Code	180612-10290-2801-03



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

**Coated Leneta Foil vs. *Listeria monocytogenes* DSM
15675**

Work Order	3109.6
Setup-Code	180612-10290-2801-03

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3109.6

Test object: Coated Leneta Foil vs. Listeria monocytogenes DSM 15675

Sample description: Coated Leneta Foil

Date of receipt of sample: 14.09.2018

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and Efficacy

Test Germ: Listeria monocytogenes DSM15675

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 180612-10290-2801-03

Sample material: Leneta Foil

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 14.9.2018
Recipient: Nano-Care Deutschland AG

Laboratory Director:

Harald Gerauer, Laboratory Director
QualityLabs BT GmbH

Released:

Markus Zehe, Managing Director
QualityLabs BT GmbH

Work Order	3109.6
Setup-Code	180612-10290-2801-03

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3109.6
Setup-Code	180612-10290-2801-03

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3109.6
Setup-Code	180612-10290-2801-03

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning	Isopropanol	-

References to deviations, preincubations, special test conditions

NONE

Work Order	3109.6
Setup-Code	180612-10290-2801-03

Test Results

	Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1	Leneta-Folie P121-10 (Reference)	102900806180001	1,3 x 10 ⁵	1,1 x 10 ⁵	1,4 x 10 ⁵	1,7 x 10 ⁴	2,9 x 10 ⁴	2,0 x 10 ⁵	-	-
2	Liquid Guard clean + primer + wipe	102900806180002				< 1,0 x 10 ¹	< 1,0 x 10 ¹	< 1,0 x 10 ¹	> 99,99	> 4
3	Liquid Guard pro +wipe	102900806180003				< 1,0 x 10 ¹	< 1,0 x 10 ¹	< 1,0 x 10 ¹	> 99,99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Listeria monocytogenes</i> DSM15675
Initial cell count inoculum / cm ²	1.25 x 10 ⁴
Initials of the editor	OS
Measurement ended on	18.06.2018

Work Order	3109.6
Setup-Code	180612-10290-2801-03

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Shendi _____

Crosschecked: Mr. Zehe _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3212.1
Setup-Code	181029-10290-2801-01



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

Coated Leneta-Foil vs. Pseudomonas aeruginosa EDCC 5272

Work Order	3212.1
Setup-Code	181029-10290-2801-01

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3212.1

Test object: Coated Leneta-Foil vs. *Pseudomonas aeruginosa* EDCC 5272

Sample description: coated foil

Date of receipt of sample: 2018-Oct-22

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: *Pseudomonas aeruginosa* EDCC 5272

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 181029-10290-2801-01

Sample material: n.b.

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 2018-Nov-07
Recipient: Nano-Care Deutschland GmbH

Laboratory Director:

 Harald Gerauer, Laboratory Director
 QualityLabs BT GmbH

Released:

 Markus Zehe, Managing Director
 QualityLabs BT GmbH

Work Order	3212.1
Setup-Code	181029-10290-2801-01

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3212.1
Setup-Code	181029-10290-2801-01

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3212.1
Setup-Code	181029-10290-2801-01

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	300	μl
Sample cleaning		-

References to deviations, preincubations, special test conditions

NONE

Work Order	3212.1
Setup-Code	181029-10290-2801-01

Test Results

Sample Name		Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1	Reference	102902310180011	5.3 x 10 ⁴	4.9 x 10 ⁴	4.2 x 10 ⁴	4.1 x 10 ⁵	3.5 x 10 ⁵	4.7 x 10 ⁵		Reference
2	Liquid Guard 2,9%	102902310180012				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Pseudomonas aeruginosa</i> EDCC 5272
Initial cell count inoculum / cm ²	1.25 x 10 ⁴
Initials of the editor	MZ
Measurement ended on	Nov-11-2018

Work Order	3212.1
Setup-Code	181029-10290-2801-01

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Zehe _____

Crosschecked: Mr. Mannala _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Work Order	3109.5
Setup-Code	180612-10290-2801-02



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and Effectiveness

Test Object:

Coated Leneta - Foil vs. Salmonella choleraesuis DSM11320

Work Order	3109.5
Setup-Code	180612-10290-2801-02

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3109.5

Test object: Coated Leneta Foil vs. Salmonella choleraesuis DSM11320

Sample description: Coated Leneta Foil

Date of receipt of sample: 14.09.2018

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and Efficacy

Test Germ: Salmonella choleraesuis DSM11320

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 180612-10290-2801-02

Sample material: Leneta Foil

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 14.9.2018
Recipient: Nano-Care Deutschland AG

Laboratory Director: _____
 Harald Gerauer, Laboratory Director
 QualityLabs BT GmbH

Released: _____
 Markus Zehe, Managing Director
 QualityLabs BT GmbH

Work Order	3109.5
Setup-Code	180612-10290-2801-02

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3109.5
Setup-Code	180612-10290-2801-02

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3109.5
Setup-Code	180612-10290-2801-02

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning	Isopropanol	-

References to deviations, preincubations, special test conditions

NONE

Work Order	3109.5
Setup-Code	180612-10290-2801-02

Test Results

Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1 Leneta-Foil P121-10 (Reference)	102900806180001	6,6 x 10 ⁴	5,5 x 10 ⁴	7,1 x 10 ⁴	7,8 x 10 ⁴	7,3 x 10 ⁴	1,2 x 10 ⁵	-	-
2 Liquid Guard clean + primer + wipe	102900806180002				< 1,0 x 10 ¹	< 1,0 x 10 ¹	< 1,0 x 10 ¹	> 99,99	> 4
3 Liquid Guard pro + wipe	102900806180003				< 1,0 x 10 ¹	< 1,0 x 10 ¹	< 1,0 x 10 ¹	> 99,99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Salmonella choleraesuis DSM11320</i>
Initial cell count inoculum / cm ²	1.25 x 10 ⁴
Initials of the editor	OS
Measurement ended on	15.06.2018

Work Order	3109.5
Setup-Code	180612-10290-2801-02

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Shendi _____

Crosschecked: Mr. Zehe_____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and Efficacy

Work Order	3109.2
Setup-Code	180611-10290-2801-02



Test Report

JIS Z 2801:2012 (Mod)

Antimicrobial products – Test for antimicrobial activity and efficacy

Test Object:

Coated Leneta-Foil vs. Staph. aureus DSM 21979

Work Order	3109.2
Setup-Code	180611-10290-2801-02

Report on Findings

Client: Nano-Care UK (Signo-Nanocare UK Ltd)
Address: PO Box 225, Oswestry, Shropshire, SY10 1DL, UK +44 1691 654 282

Work order no.: 3109.2

Test object: Coated Leneta-Foil vs. Staph. aureus DSM 21979

Sample description: Coated Foil

Date of receipt of sample: 08.06.2018

Type of test: JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy

Test Germ: Staphylococcus aureus DSM21979 EDCC 5247

Test laboratory: QualityLabs BT GmbH

Address: Neumeyerstrasse 46a
90411 Nuremberg, Germany

Setup-Code: 180611-10290-2801-02

Sample material: Leneta Foil

No. of pages in report: 7

Report on findings to the client: **Place and date of preparation:** Nuremberg, 13.9.2018
Recipient: Nano-Care Deutschland AG

Laboratory Director:

 Harald Gerauer, Laboratory Director
 QualityLabs BT GmbH

Released:

 Markus Zehe, Managing Director
 QualityLabs BT GmbH

Work Order	3109.2
Setup-Code	180611-10290-2801-02

Declaration on Quality Assurance

This investigation was performed and supervised according to the standard operating procedure "SOP zu JIS Z 2801:2012 (Mod)" by QualityLabs BT GmbH. The laboratory and process are continually monitored by independent, external authorities, as well as by internal audits.

Archiving

A copy of the test report, a protocol of the measurement as well as the accompanying correspondence and business records are archived by QualityLabs BT GmbH. The retention period is at least 10 years.

Test description

Anti-bacterial activity is determined in accordance with a modified version of JIS Z 2801:2012.

During the test, a thin liquid-film containing the bacteria (1.25×10^4 CFU / cm²) is applied directly to the test sample (Standard: 5 cm x 5 cm). To avoid desiccation a foil (Standard: 4cm x 4cm, Stomacher Bags) is applied. Immediately after inoculation, the bacteria from the reference sample are separated from the sample and the enveloping foil surfaces using ultrasound and vortex devices and the number of viable germs (CFU – colony-forming units) is determined (t_0 value). A further set of reference samples and samples given anti-microbial treatment is incubated with bacteria in a liquid-film and the enveloping foil in a damp environment at 37°C. After 24 hours, the bacteria are separated from the sample surfaces using ultrasound and vortex devices and the number of viable germs is determined (t_{24} value).

Work Order	3109.2
Setup-Code	180611-10290-2801-02

Assessment of antimicrobial activity

A logarithmic germ reduction of **≥ 3 log scales** of the antimicrobial sample in comparison to the respective reference is used as assessment criterion to pass the antimicrobial test.

Germ reduktion [log scales]	Antibacterial activity
< 3	Not sufficient antimicrobial activity
≥ 3	Sufficient antimicrobial activity

Work Order	3109.2
Setup-Code	180611-10290-2801-02

References to Testconditions

Testconditions		
Sample size	25	cm ²
Foil size	16	cm ²
Volume Inoculum	400	μl
Sample cleaning	Isopropanol	-

References to deviations, preincubations, special test conditions

NONE

Work Order	3109.2
Setup-Code	180611-10290-2801-02

Test Results

Sample Name	Sample Code	t ₀ (cells/cm ²)			t ₂₄ (cells/cm ²)			Reduction [%]	Log Reduction
1 Leneta-Folie P121-10 (Reference)	102900806180001	9.0 x 10 ⁴	1.1 x 10 ⁵	1.1 x 10 ⁵	5.1 x 10 ⁵	4.1 x 10 ⁵	5.3 x 10 ⁵	-	-
2 Liquid Guard clean + primer + wipe	102900806180002				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4
3 Liquid Guard pro + wipe	102900806180003				< 1.0 x 10 ¹	< 1.0 x 10 ¹	< 1.0 x 10 ¹	> 99.99	> 4

*see "Interpretation of Results", page 6

Test strain	<i>Staphylococcus aureus</i> DSM21979 EDCC 5247
Initial cell count inoculum / cm ²	1.25 x 10 ⁴
Initials of the editor	MZ
Measurement ended on	Mar-12-20xx

Work Order	3109.2
Setup-Code	180611-10290-2801-02

Comments on test objects

NONE

Interpretation of the results based on the measurements

NONE

Editor: Mr. Zehe _____

Crosschecked: Mr. Shendi _____

References

JIS Z 2801:2012 Antimicrobial products – Test for antimicrobial activity and efficacy